

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 22 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130185</u> ✓
Name <u>BRANDON LAWRENSEN</u>		ID# <u>66609</u> Date <u>06/12/09</u> ✓
A	Agency <u>JUNEAU POLICE DEPT</u>	Phone # <u>586-0600</u>
Instrument Location <u>JUNEAU POLICE DEPT</u> ✓		
Alco S/N <u>77675</u> ✓		Target Value <u>.080</u> ✓ High Pressure <u>400</u>
B	Alco Test Values	
	<u>.085</u> ✓ 1 <sup>st</sup> Alco	<u>.084</u> ✓ 2 <sup>nd</sup> Alco
Signature <u>[Signature]</u> ✓		<u>86/12/09</u> <u>MB</u> <u>955/9</u>
(OVER)		

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 26<sup>th</sup> day of June 2009.

[Signature] (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public; State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130185 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130185

JUNE 12, 2009 ✓

OPERATOR'S NAME:

LAWRENSON/BRANDON/J

OPERATOR'S NUMBER: 6609

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 123456

DEPT/AGENCY: JNU1

CASE/REPORT: 0123456

TEST TYPE: V

ALCO TARGET VALUE: .080 ✓

ALCO S/N: 77675 ✓

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 30.14 in		
ALCO TARGET	.080	09:51
BLANK TEST	.000	09:51
INTERNAL STANDARD	VERIFIED	09:51
ALCO TV 30.14 in	.085	09:52
BLANK TEST	.000	09:53
SUBJECT SAMPLE	.000	09:53
BLANK TEST	.000	09:54
ALCO TV 30.14 in	.084	09:54
BLANK TEST	.000	09:55

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130185

JUNE 12, 2009 ✓

TIME 09:49

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 41c

BAROMETER: 30.14 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ \* -